

CALIBRATION CERTIFICATE

Customer Details-

Robert Hopkins Environmental Ltd
Bullock Street
West Bromwich
West Midlands
B70 7HE

Cal Cert Ref-	T16310556-02/03/2022
---------------	----------------------

Scale Details		As Found Linearity & Hysteresis		
Location	Site	LOAD	READING	DEVIATION
Serial Number	T16310556	0kg	0kg	0
Manufacturer	Vishay VT300	4000kg	4000kg	0
Capacity	40000kg x 10kg	8000kg	8000kg	0
Machine Type	Weighbridge	12000kg	12000kg	0
Calibration Date	02/03/2022	16000kg	16000kg	0
Next Calibration Due	March 2023	20000kg	20000kg	0
Tolerance	+/- 3 Divisions	24240kg	24240kg	0

REPEATABILITY				
LOAD	READING 1	READING 2	READING 3	READING 4
24240kg	24240kg	24240kg	24240kg	24240kg

ECCENTRIC LOADING							
POSITION	LOAD	READING	INDICATOR				
1	6240kg	6240kg	<table border="1"> <tr> <td>1</td> <td>2</td> </tr> <tr> <td>3</td> <td>4</td> </tr> </table>	1	2	3	4
1	2						
3	4						
2	6240kg	6240kg					
3	6240kg	6260kg					
4	6240kg	6240kg					

Compliance Weights Used		Calibration of above machine	
Box Set No	Cal Cert No	Tested By-	Jay Hatton
PM51-75, 47-48, 1-2	W22010/W22011	Signature-	<i>J.Hatton</i>
All test weights are calibrated and traceable to National Standards.			
Make up weights are also used FLT 4240kg			